

# DEMKO CERTIFICATE

<b>Certificate No.</b>	D-07677
<b>Page</b>	1/3
<b>Date of Issue</b>	2019-12-30
<b>Certificate Holder</b>	SL POWER ELECTRONICS CORP BLDG A 6050 KING DR VENTURA, CA 93003 USA
<b>Manufacturer</b>	SL POWER ELECTRONICS CORP BLDG A 6050 KING DR VENTURA, CA 93003 USA
<b>Production site</b>	INDUSTRIAS S L S A DE C V CIRCUITO SIGLO XXI 2055 COL PARQUE INDUSTRIAL EX-XXI MEXICALI, BC 21254 Mexico  See Page 2
<b>Certified Product Model</b>	Power supply NGB425SXXYZZ See Page 2
<b>Trademark</b>	 See Page 2
<b>Rated Voltage / Frequency</b>	See Page 2
<b>Rated Current / Power</b>	See Page 2
<b>Insulation Class</b>	Class I or II
<b>Degree of protection (IP)</b>	--
<b>Tested acc. to</b>	EN 60601-1:2006/A12:2014, EN 60601-1:2006/A1:2013, EN 60601-1:2006
<b>Test Report No.</b>	E511446-D1001-1/A0/C0-CB issued on 2019-12-26
<b>Additional</b>	Component for built-in. The Risk Management requirements of the standard were not addressed in the report.
<b>Expire date</b>	2029-12-29

**Certification Manager**  
Jan-Erik Storgaard

This is to certify that representative sample(s) of the Product described herein ("Certified Product") have been investigated and found in compliance with the Standard(s) indicated on this Certificate, in accordance with the D Mark Requirements. The Designated Certificate holder is entitled to use the d or for cables <DEMKO> for the Certified Product manufactured at the production site(s) identified above, in accordance with the D Mark Service Agreement, including without limitation the D Mark Testing and Certification Services Service Terms. Only those Products bearing the D Mark should be considered as being covered by UL's D Mark Service. This Certificate shall remain valid through the expiration date, unless terminated earlier in accordance with the Service Agreement including without limitation if the Standard identified on this Certificate is amended or withdrawn prior to the expiration date.

**Certification Body**

UL International Demko A/S, Borupvang 5A, DK-2750  
Ballerup, Denmark, Tel. +45 44 85 65 65,  
info.dk@ul.com  
www.ul-europe.com



# Appendix DEMKO CERTIFICATE

**Certificate No.** D-07677  
**Page** 3/3  
**Date of Issue** 2019-12-30

## Model Details:

NGB425SXXYZZ Where XX represents the output voltage which may be any number from 12 to 48. Y can be K (for Class I) or C (for Class II); ZZ can be any number between 00-99, blank or any letter from AA to ZZ, only for market purpose, not influence safety function.

## Production Sites:

SL XIANGHE POWER ELECTRONICS CORP  
No.B-02-03,North side of Landscape Ave, Qibu District,  
Environmental Industrial Park  
Xianghe, 065400 Hebei China

## Ratings:

Input: 100-240 Vac, 50-60 Hz, 5.5A

### Output:

Model NGB425S12K and NGB425S12C

### For convection:

Output1: 12Vdc/19.5A

Output2: 5Vdc/0.5A

### For 400LFM:

Output1: 12Vdc/30.0A

Output2: 5Vdc/1.0A

Model NGB425S15K and NGB425S15C

### For convection:

Output1: 15Vdc/13.17A

Output2: 5Vdc/0.5A

### For 400LFM:

Output1: 15Vdc/24.0A

Output2: 5Vdc/1.0A

Model NGB425S24K and NGB425S24C

### For convection:

Output1: 24Vdc/11.15A

Output2: 5Vdc/0.5A

### For 400LFM:

Output1: 24V/17.5A

Output2: 5Vdc/1.0A

Model NGB425S48K and NGB425S48C

### For convection:

Output1: 48Vdc/5.57A

Output2: 5Vdc/0.5A

### For 400LFM:

Output1: 48Vdc/8.75A

Output2: 5Vdc/1.0


**Certification Body**

UL International Demko A/S, Borupvang 5A, DK-2750  
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@ul.com  
www.ul-europe.com



# Appendix DEMKO CERTIFICATE

Certificate No. D-07677  
Page 3/3  
Date of Issue 2019-12-30



---

**Certification Body**

UL International Demko A/S, Borupvang 5A, DK-2750  
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@ul.com  
www.ul-europe.com

