



Excelsys Technologies Ltd.
27 Eastgate Drive
Eastgate Business Park
Little Island
Co. Cork, Ireland

Tel: + 353 21 4354716
Fax: + 353 21 4354864

DECLARATION OF CONFORMITY

Product Name: UltiMod Series

Product Type: Component Power Supply

Model Number: UXz abcdef g k h j for 6 slot UltiMod powerPac units populated with powerMod modules or
UXz abcd g k h j for 4 slot UltiMod powerPac units populated with powerMod modules

Where:

"UX" = All UltiMod part numbers start with "UX"

"z" = 6, 4, D, S or X

6 = 6 slot UltiMod powerPac

4 = 4 slot UltiMod powerPac

D = 6 slot UltiMod powerPac dual input fused

S = 4 slot UltiMod powerPac with reinforced isolation to earth on outputs a and b

X = 6 slot UltiMod powerPac limited to 900W

"a" = 0, 1, 2, 3, 4, 5, 6, 7, 8, A, B, C, D, E, F, G, H, J, K, L, M, N, P, Q, R or T

"b" = 0, 1, 2, 3, 4, 5, 6, 7, 8, A, B, C, D, E, F, G, H, J, K, L, M, N, P, Q, R or T

"c" = 0, 1, 2, 3, 4, 5, 6, 7, 8, A, B, C, D, E, F, G, H, J, K, L, M, N, P, Q, R or T

"d" = 0, 1, 2, 3, 4, 5, 6, 7, 8, A, B, C, D, E, F, G, H, J, K, L, M, N, P, Q, R or T

"e" = 0, 1, 2, 3, 4, 5, 6, 7, 8, A, B, C, D, E, F, G, H, J, K, L, M, N, P, Q, R or T

"f" = 0, 1, 2, 3, 4, 5, 6, 7, 8, A, B, C, D, E, F, G, H, J, K, L, M, N, P, Q, R or T

"g" = '-', P, C, R or S

'-' = Standard model (nominal voltage)

P = Specific voltage adjustment settings

C = Conformal coating

R = Ruggedised for vibration

S = C + R

"k" = Any alphanumeric character describing customer internal wiring lengths. Where no internal wiring exists and standard IEC appliance inlet is used k=0.

"h" = 0, 2, 4 or 6

0 = Standard model

2 = Reverse fan

4 = Low leakage

6 = 2 + 4

"j" = Any alphanumeric character. Optional. Logistics use only.

See Page 2.

Excelsys Technologies Ltd. confirms that this product conforms to:

IEC 60601-1-2: 2014 (4th Edition)

Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests

Testing was conducted by Compliance Engineering Ireland Ltd.

Place of Issue: Cork, Ireland
Date of Issue: 1st July 2016

For and on behalf of
Excelsys Technologies Limited



Dermot Flynn