



Ref. Certif. No.

**US-38490-UL**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

**CB TEST CERTIFICATE**

Product	Medical Power Supply
Name and address of the applicant	SL POWER ELECTRONICS CORP BLDG A 6050 KING DR VENTURA CA 93003 USA
Name and address of the manufacturer	SL POWER ELECTRONICS CORP BLDG A 6050 KING DR VENTURA CA 93003 USA
Name and address of the factory	SL XIANGHE POWER ELECTRONICS CORP No.B-02-03, North side of Landscape Ave, Qibu Distr Environmental Industrial Park Xianghe,Hebei 065400 China
Note: When more than one factory, please report on page 2	<input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
Ratings and principal characteristics	Input: 100-240Vac, 50-60Hz, 4.0A <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
Trademark (if any)	N/A
Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	SLB300SXXYZZ <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
Additional information (if necessary may also be reported on page 2)	The risk management requirements of the standard were not addressed. For Class II. <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005
As shown in the Test Report Ref. No. which forms part of this Certificate	E511446-D1022-1/A0/C0-CB issued on 2021-06-30

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2021-07-20

Signature: Jolanta M. Wroblewska



Ref. Certif. No.

**US-38490-UL**

**Factories:**

INDUSTRIAS S L S A DE C V  
CIRCUITO SIGLO XXI 2055 COL PARQUE INDUSTRIAL EX-XXI 21254 MEXICALI BC  
Mexico

**Additional Model(s):**

**Series:** SLB300SXXYZZ, Where XX represents the output voltage which may be any number from 12 to 56. Y can be C (for Class II); ZZ can be any number between 00-99, blank or any letter from AA to ZZ, only for market purpose, not influence safety function.

**Additional Ratings:**

Output: 12-56Vdc, 22.5-5.36A, 300W max.  
Standard models output rating are listed as below:  
For SLB300S12YZZ: 12Vdc/22.5A, Max. 270W  
For SLB300S24YZZ: 24Vdc/12.5A, Max. 300W  
For SLB300S48YZZ: 48Vdc/6.25A, Max. 300W  
For SLB300S56YZZ: 56Vdc/5.36A, Max. 300W

**Additionally evaluated to:**

EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014  
National Differences specified in the CB Test Report.

**Additional information (if necessary)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Signature:

Jolanta M. Wroblewska

Date: 2021-07-20